EU DECLARATION OF CONFORMITY

CE 0482

We, TELEMED UAB

located at Highway Business Centre

Savanoriu pr. 178A Vilnius LT-03154 Lithuania

Declare under our sole responsibility that:

Products: Ultrasound scanner Micrus EXT-1H

Ultrasound Transducers C5-2R60S-3 MC4-2R20S-3

L12-5L40S-3 MC8-4R20S-3 L12-5N40-M3 MC10-5R10S-3 L12-5N40-MV3 MCV9-5R10S-3

L15-6L25S-3

Classification: Class IIa (in compliance with Annex II, Art.11 Medical Device Directive) are in conformity with:

Essential Requirements of Council Directive 93/42/EEC (Medical Device Directive)

EN 60601-1:2006, EN 60601-1:2006/AC:2010 Medical electrical equipment – Part 1: General requirements for basic safety and essential performance (IEC 60601-1: 2005)

EN 60601-1-2:2015 Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests (IEC 60601-1-2:2014)

EN 60601-2-37:2008, EN 60601-2-37:2008/A11:2011, EN 60601-2-37:2008/A1:2015 Medical electrical equipment – Part 2-37: Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment (IEC 60601-2-37:2007, IEC 60601-2-37:2007/AMD1:2015)

EN ISO 10993-1:2009 Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2009)

EN ISO 10993-5:2009 Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity (ISO 10993-5:2009)

EN ISO 10993-10:2010 Biological evaluation of medical devices – Part 10: Tests for irritation and skin sensitization (ISO 10993-10:2010)

EN 62304:2006 Medical device software – Software life cycle processes (IEC 62304:2006)

EN ISO 14971:2019 Medical devices – Application of risk management to medical devices (ISO 14971:2019)

ISO 15223-1:2021 Medical devices - Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.

In addition, we ensure and declare that the distributed products do not contain a medicinal substances or materials derived from animal or human tissue.

The compliance with the Council Directive MDD 93/42/EEC is under the monitoring of the Notified Body: **MEDCERT GmbH, Pilatuspool 2, 20355 Hamburg, code: 0482**

Vilnius, October 4, 2021

Dmitry Novikov, President